

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
STC Solutions Ltd Unit 1 Cygnus Orion Court Great Blakenham Ipswich IP6 0LW CONTACT  Peter Ord	MAKE & MODEL System 2X & 1530  SERIAL NO. 123456 CAPACITY 300kg DIVISION 0.1kg LOCATION Workshop  I.D. TOLERANCE +/- 3 divisions	CERT NO. 559691A TESTED BY Aaron Bailey DATE TESTED February 2017 NEXT TEST DUE February 2018	SET NO. / CERT NO. STC: UF0816/UF1300R STC: UE1782/UF0069

**AS FOUND**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg	300.0kg	306.0kg	+ 6.0kg
10.0kg	10.2kg	+ 0.2kg			
20.0kg	20.4kg	+ 0.4kg			
50.0kg	51.0kg	+ 1.0kg			
100.0kg	102.0kg	+ 2.0kg			
200.0kg	204.0kg	+ 4.0kg			

**DEFINITIVE**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg	300.0kg	300.0kg	0.0kg
10.0kg	10.0kg	0.0kg			
20.0kg	20.0kg	0.0kg			
50.0kg	50.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			


**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	50.0kg	51.0kg	+ 1.0kg
2	50.0kg	51.0kg	+ 1.0kg
3	50.0kg	51.0kg	+ 1.0kg
4	50.0kg	51.0kg	+ 1.0kg

**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	50.0kg	50.0kg	0.0kg
2	50.0kg	50.0kg	0.0kg
3	50.0kg	50.0kg	0.0kg
4	50.0kg	50.0kg	0.0kg

**LAYOUT**



**COMMENTS OR OBSERVATIONS**

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord      Position. Service Manager      Signature. 