

Calibration Certificate Test Record

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
STC HIRE STOCK 1 Cygnus Orion Court Great Blakenham Ipswich Suffolk IP6 0LW <u>CONTACT</u> Clark Skates	<u>MAKE & MODEL</u> Ohaus Ranger Count <u>SERIAL NO.</u> 3230809 <u>CAPACITY</u> 6000g <u>DIVISION</u> 0.1g <u>LOCATION</u> <u>I.D</u> TOLERANCE +/- 3 divisions	<u>CERT NO.</u> 575855H <u>TESTED BY</u> Aaron Bailey <u>DATE TESTED</u> September 2022 <u>NEXT TEST DUE</u> September 2023	<u>SET NO. / CERT NO.</u> BW3: UM0353

AS FOUND

LOAD POINT	READING	ERROR + / -
0.0g	0.0g	0.0g
1000.0g	1000.0g	0.0g
2000.0g	1998.0g	- 2.0g
3000.0g	2996.0g	- 4.0g
4000.0g	3994.0g	- 6.0g
5000.0g	4992.0g	- 8.0g
6000.0g	5990.0g	- 10.0g

DEFINITIVE

LOAD POINT	READING	ERROR + / -
0.0g	0.0g	0.0g
1000.0g	1000.0g	0.0g
2000.0g	2000.0g	0.0g
3000.0g	3000.0g	0.0g
4000.0g	4000.0g	0.0g
5000.0g	5000.0g	0.0g
6000.0g	6000.0g	0.0g

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR + / -
N/A			

CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR + / -
N/A			

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.**